



National Accreditation Board for
Testing and Calibration Laboratories

CERTIFICATE OF ACCREDITATION

HI-TECH CALIBRATION & TESTING LLP

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

**"General Requirements for the Competence of Testing &
Calibration Laboratories"**

for its facilities at

GALA NO.: 60, ROYAL INDUSTRIAL HUB, VILLAGE: VALWADA, UMBERGAON, VALSAD, GUJARAT,
INDIA

in the field of

CALIBRATION

Certificate Number: CC-2478

Issue Date: 14/12/2023

Valid Until: 03/01/2025

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.
(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Entity: HI-TECH CALIBRATION & TESTING LLP

Signed for and on behalf of NABL



N. Venkateswaran
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

HI-TECH CALIBRATION & TESTING LLP, GALA NO.: 60, ROYAL INDUSTRIAL HUB, VILLAGE: VALWADA, UMBERGAON, VALSAD, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2478

Page No

17 of 63

Validity

14/12/2023 to 03/01/2025

Last Amended on

-

AT LAB

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
95	FLUID FLOW-FLOW MEASURING DEVICES	Air Velocity / Anemometer / Pitot Tube / Flow Velocity sensor with Indicator	Using Wind Tunnel & Pitot Tube or Anemometer by Comparison method	3 m/s to 30 m/s	4.3%
96	FLUID FLOW-FLOW MEASURING DEVICES	Flow meter/Air Flow meter/Rota meter/Gas Flow meter	Using Digital Flow Calibrator By Comparison Method	30 cc/min to 3000 cc/min	2.72%
97	FLUID FLOW-FLOW MEASURING DEVICES	Flow meter/Air Flow meter/Rotameter/Gas Flow meter Air Flow Calibrator	Using Digital Flow Calibrator by Comparison method	1 lpm to 100 lpm	2.0%
98	FLUID FLOW-FLOW MEASURING DEVICES	Respirable Dust Sample/Gas Flow/PM 1.0 or PM 2.5 Sampler	Using Top Loading Calibrator By Comparison Method	0.250 cu. m/min to 1.5 cu. m/min	2.5%

ON SITE

35	FLUID FLOW-FLOW MEASURING DEVICES	Flow meter/Air Flow meter/Rota meter/Gas Flow meter	Using Digital Flow Calibrator By Comparison Method	30 cc/min to 3000 cc/min	2.72%
36	FLUID FLOW-FLOW MEASURING DEVICES	Flow meter/Air Flow meter/Rotameter/Gas Flow meter Air Flow Calibrator	Using Digital Flow Calibrator by Comparison method	1 lpm to 100 lpm	2.0%